

Produktänderungsinformation

Ausgabedatum: 2024-05
Betroffene Artikel: PM1026-T
Beschreibung der Änderung: Schaltungsanpassung wegen Bauteilabkündigung

Parameter	alt	neu
Messfehler TRMS (200 mV)	$\pm 0,5\%$ v. Messwert ± 1 digit, crest ≤ 5 , $f < 1$ kHz	$\pm 0,5\%$ vom Messwert ± 1 digit bei DC, 50 Hz 800 Hz und crest $< 1,5$; bzw. $\pm 1\%$ vom Messwert bei 6 kHz
Nullpunktoffset TRMS		+25 mV / +50 μ A am Analogausgang

Änderung Datenblatt: 2024-05
Lieferung ab: 2024-05
Identifikation: Lieferdatum nach 01.05.2024

Product Change Notification

Issue date: 2024-05
Affected products: PM1026-T
Object of change: Circuit adjustment due to component discontinuation

Parameter	old	new
Total error TRMS (200 mV)	± 0.5 % of value ± 1 count, crest ≤ 5 , $f < 1$ kHz	± 0.5 % of measured value ± 1 digit at DC, 50 Hz ... 800 Hz and crest < 1.5 ; resp. ± 1 % of measured value at 6 kHz
Zero-point offset		+25 mV / +50 μ A on analog output

Change of data sheet: 2024-05
Effective from: 2024-05
Identification: supply voltage on device label